

This Page Is Inserted by IFW Operations
and is not a part of the Official Record

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images may include (but are not limited to):

- BLACK BORDERS
- TEXT CUT OFF AT TOP, BOTTOM OR SIDES
- FADED TEXT
- ILLEGIBLE TEXT
- SKEWED/SLANTED IMAGES
- COLORED PHOTOS
- BLACK OR VERY BLACK AND WHITE DARK PHOTOS
- GRAY SCALE DOCUMENTS

IMAGES ARE BEST AVAILABLE COPY.

**As rescanning documents *will not* correct images,
please do not report the images to the
Image Problem Mailbox.**




INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification 6 : A61B 17/17	A1	(11) International Publication Number: WO 97/13467 (43) International Publication Date: 17 April 1997 (17.04.97)
(21) International Application Number: PCT/IT96/00179 (22) International Filing Date: 24 September 1996 (24.09.96) (30) Priority Data: BO95U000144 13 October 1995 (13.10.95) IT (71)(72) Applicant and Inventor: CATAMO, Lucio [IT/IT]; Via Montencero, 26, I-40131 Bologna (IT).	(81) Designated States: AU, CA, CN, JP, US, European patent (AT, BE, CH, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE). Published With international search report. Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.	

(54) Title: PROBE FOR LOCATING HOLES IN ORTHOPAEDIC INTERNAL FIXATION DEVICES

(57) Abstract

The necessity of the use, sometimes excessive, of an image intensifier in order to localize the holes for the screws in the internal fixation procedure with an interlocking nail implies an increase of the risks of radiation exposure involving the orthopaedic surgeon, the operation theatre personnel and the patient. The use is proposed of a magneto-resisting probe which looks for a signal produced by 2 microelectromagnets fixed to a rod which is introduced along the intramedullary nail: a high frequency alternating signal is produced by the two microelectromagnets and is recorded by the probe on the skin. The probe is made of 4 magneto-resistive sensors arranged on orthogonal axes for a more precise and straightforward location in the x and y directions. A screen with 4 arrows shows the direction one should follow for correct centering over the holes. A fourth possibility is represented by a small neodymium magnet, which can be inserted as a plug in the nail hole or fixed to a rod inserted into the canal of the nail, allowing a quick and precise localization of the hole and reducing the exposition time to radiation as well as the overall surgical procedure duration.



FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AM	Armenia	GB	United Kingdom	MW	Malawi
AT	Austria	GE	Georgia	MX	Mexico
AU	Australia	GN	Guinea	NE	Niger
BB	Barbados	GR	Greece	NL	Netherlands
BE	Belgium	HU	Hungary	NO	Norway
BF	Burkina Faso	IE	Ireland	NZ	New Zealand
BG	Bulgaria	IT	Italy	PL	Poland
BJ	Benin	JP	Japan	PT	Portugal
BR	Brazil	KE	Kenya	RO	Romania
BY	Belarus	KG	Kyrgyzstan	RU	Russian Federation
CA	Canada	KP	Democratic People's Republic of Korea	SD	Sudan
CF	Central African Republic	KR	Republic of Korea	SE	Sweden
CG	Congo	KZ	Kazakhstan	SG	Singapore
CH	Switzerland	LI	Liechtenstein	SI	Slovenia
CI	Côte d'Ivoire	LK	Sri Lanka	SK	Slovakia
CM	Cameroon	LR	Liberia	SN	Senegal
CN	China	LT	Lithuania	SZ	Swaziland
CS	Czechoslovakia	LU	Luxembourg	TD	Chad
CZ	Czech Republic	LV	Latvia	TG	Togo
DE	Germany	MC	Monaco	TJ	Tajikistan
DK	Denmark	MD	Republic of Moldova	TT	Trinidad and Tobago
EE	Estonia	MG	Madagascar	UA	Ukraine
ES	Spain	ML	Mali	UG	Uganda
FI	Finland	MN	Mongolia	US	United States of America
FR	France	MR	Mauritania	UZ	Uzbekistan
GA	Gabon			VN	Viet Nam

1 TITLE

PROBE FOR LOCATING HOLES IN ORTHOPAEDIC INTERNAL FIXATION DEVICES

TECHNICAL FIELD

- 2 Many different techniques are employed in current orthopaedic and traumatologic practice according to the bone segment involved and to special surgical indications. In long bone fractures the use of intramedullary metallic means, static, dynamic or mixed (Grosse Kempf, Kuntscher, A.O., etc.), control segmental rotation and maintain the desired limb length. The great advantage of intramedullary fixation is
- 3 demonstrated by the simple technique, by the reduced extension of the skin incision for the nail insertion and by the prevention of fracture site opening. Nevertheless such procedure requires repeated x-ray controls, or image intensifier shots, with exposure of the patient and the O.R. personnel to ionizing radiations.

BACKGROUND ART

- 4 These controls are necessary to guide the nail into the medullary canal of the second fracture segment and subsequently to fix the screws which interlock the nail and neutralize the rotation forces keeping the limb length. Despite the validity of the different centering gauges, the amplioscope is frequently and excessively used to recognize the nail holes, mainly the distal ones, and to fit the corresponding
- 5 screws in, increasing in this way the radiological risks.

The enthusiasm for this fixation technique due to its simplicity and undeniable validity has been cooled down by the excessive repeated exposure of the orthopaedic surgeon to ionizing radiations.

- 6 The proposed solutions to accomplish this operation, although functional and simple "in vitro", have showed their limits "in vivo" where the reduction of radiological exams in order to center the interlocking screw into the corresponding hole was not satisfactory.

- 7 Many orthopaedic surgeons have therefore aimed their interest to other fixation systems, less valid and with a more limited indication range, but which did not demand a frequent use of radiological pictures.

DISCLOSURE OF INVENTION

The recognition of the hole of the intramedullary nail and the subsequent aiming of a screw is possible by a simple and harmless device which uses an electromagnetic probe that localizes metallic means and reveals the holes with on-off mechanism.

- 8 The probe is made by an inductive sensor of proximity which an internal high frequency oscillator that produces an electromagnetic field in the immediate vicinity of the sensor. The presence of a metallic object (actioner) in the action field of the sensor reduces the wideness of the oscillation since part of the electromagnetic energy is transferred from the sensor to the actioner on which dissipates for the effect of the parasitic currents of Foucault. The wideness of the oscillation then decreases with the distance between actioner and sensor and may bring analogic information on the position of the object at the exit of the sensor or may be converted into a digital signal on-off by means of a Schmitt trigger. An electric circuit transforms this message into an acoustic and/or light (led) signal. In the presence of a metallic intramedullary nail the probe sensitizes and shows the presence with the acoustic and/or light signal. When the probe passes above the hole the signal stops. A microprocessor device converts this on/off signals on a monitor which allows a real time identification of the object and its configuration (hole: off).
- 11 The probe consists of a head (site of the electromagnetic device) and a handle which is guided by the operator sliding along the limb where a nail is present. The overall length is 15-20 cm. The head has a circular configuration and delimits a central hole; while sliding along the limb indicates the presence of the metallic object; at the level of the nail hole the signal fades till a complete interruption once the probe hole overlaps the nail hole. The skin point where the probe hole lies is the exact point where the incision may be made to drive correctly the interlocking screw.

- Alternatively, one can use an electromagnetic probe with a sensor (sensor of Hall) which reveals the electromagnetic field produced by a small magnet. This small magnet is inserted in the hole of the nail inside a small pull-out screw (of metal or teflon). The probe contains a stabilized feeder, an hall-effect sensor, a differential amplifier, a Schmitt trigger, a low power transistor with open collector. In resting condition the base of the final transistor is not polarized therefore that transistor does not conduce. If, on the other hand, one approaches the south pole of a magnet to the body of the sensor, or the north pole to the opposite site, automatically one stimulates the Hall-effect sensor which on its turn generates a weak tension with an intensity proportional to the one of the applied magnetic field. This tension, amplified by the "differential", is applied at the entrance of the Schmitt trigger and, as soon as its value overtakes the threshold level of that trigger, the latter will produce a positive tension upon the base of the final transistor and will bring it into conduction.

- As third possibility, one can use a magneto-resistive sensor made of special "bridge" resistances with a component sensitive to variations of the magnetic field which crosses them, in order to change the inner conductivity according to such field. Having a bridge connection and feeding them with a continuous tension on two opposite sides the bridge, one can obtain a variation of the differential tension

on the other two sides proportional to the variation of the crossed magnetic field. The magnet which stimulates probe is characterized by a tiny cylindrical magnet of few millimeters, able to produce a magnetic field axial to the intramedullary nail hole. It is inserted in the head of a screw or a rod (metallic, teflon, PVC or other material) which temporarily plugs the nail hole. Once the tiny magnet is applied to the nail the magneto-resistive probe records the magnetic field which corresponds to the hole.

A fourth possibility is represented by the electromagnetic signal produced by 2 microelectromagnets fixed to the rod which is introduced along the intramedullary nail: a high frequency signal, between 20 KHz and 1 Mhz, is produced alternatively by the two microelectromagnets and is recorded by the probe on the skin. The probe is made of 4 magneto-resistive sensor in crossed position for a more precise and straightforward research in the north, south, west and east directions. A shielded wire brings the message to an electric circuit set at the desired distance which transforms this message into an acoustic or light signal. An alphanumeric system allows to quantify the distance. A screen with 4 arrows (above, below, right and left) shows the direction one should follow for a correct centering. A special device connected to the probe enables the sensitivity selfcorrection: it is not necessary therefore to manually correct the powermeter and one can limit the research to few essential gestures. The probe is contained in a small sterilisable cylinder with the shape of a pen, a few millimeters of diameter, and allows the longitudinal gliding along the limb which hosts the nail. It is introduced in a specific guide which, once the hole has been localized, allows the extraction of the probe and the introduction of the drill which pierces the bone without losing the precise alignment with the hole.

The screw containing the small magnet that plugs the nail hole will then be extracted with a screwdriver, the whole rod containing the magnet will be directly pulled out or a metallic rod will be used to catch the magnet.

22 BRIEF DESCRIPTION OF DRAWINGS

Fig. 1. Electromagnetic probe Iron S3

Fig. 2. Electromagnetic probe for the recognition of metallic bodies and the discrimination of their holes during special internal fixation techniques Orthopaedics.

23 Fig. 3. Two microelectromagnets emitting coils in a metallic or teflon rod (with preprinted circuit) in the medullary nail.

Fig. 4, 5, 6. Intramedullary metallic means in long bone fractures. The coil emits an electromagnetic field with the main magnetic axis corresponding to the hole axis. Blocking system of the rod.

- 24 Fig. 7. In long bone fractures the use of intramedullary metallic means. The head has a circular configuration and delimits a central hole. The 4 (8) magnetic sensor have their sensitive surfaces a plane at 90° to the tube axis. The guide-tube allows the way through of the drill for the bone hole without losing the correct adjustment against the cortical bone, thanks to the special conformation of the head.
- 25 Fig 8. Monitor. On the graphic display unit the corresponding cartesian references are visualized.
- Fig. 9. Iron S3. Generator - blocks.
- Fig. 10. Circuit.
- Fig. 11. Probe.

26 BEST MODE FOR CARRYING OUT THE INVENTION

- A current generator at 20 Khz sinusoidal frequency, with a rough 200 mA RMS intensity, is alternatively connected to 2 emitting coils in a mettalic or teflon rod (with preprinted circuit) in the medullary nail. The coil emits an electromagnetic field with the main magnetic axis corresponding to the hole axis.
- 27 The detecting probe is formed by a teflon handle which supports an holed cylinder, containing 4 coils, each on a magnetic iron "semiolla", with the open side looking toward the head of the probe. The metallic guide-tube is inserted in the cylinder hole with the faceted and pointed extremity for the adjustment against the cortical bone.
- 28 The 4 magnetic iron sensor have their sensitive surfaces on a plane at 90° to the tube axis, with at the final phase of the research coincides whit the one of the searched magnetic field.
- The positions on the plane are simmetric and equidistant (crosswise). When the probe is on the searched axis the signals induced by the field itself are equidistant from the axis and therefore have equivalent intensity.
- 29 Such signal are conditioned, filtered and amplified in tension and current by a superficially mounted circuit inserted into the probe handle. The 4 amplified signal are transmitted by the connection cable to the instrumend containig the generator where they are received, refiltered and reamplified. Each signal is then revealed in
- 30 wideness by a dedicated circuit and transformed into continous current by a trimmer of final amplification adjustment.
- The signal in continuos current, with a tension equivalent to the wideness of the corresponding in alternating current, are measured by a microcomputer containig a digital analogic converter.
- 31 The obtained measurements are again filtered at software level, visualized on a graphic and alphanumeric display unit. It is then obtained the difference of numeric values of each signal couple, relative to a diagonal of the cross. The ones relative to the diagonal parallel to the medullary rod form the X axis; those at 90° form the Y axis.

- 32 On the graphic display unit the corresponding cartesian references are visualized. If the sum of the 4 monitored signals overtakes a minimal predefinite value a light indicator on the cartesian plane is displayed, in order to obtain the indicator in the center when the 4 signal are equivalent, or on the relative axis of the single couples with $\Delta=0$.
- 33 The guide-tube allows the way through of the drill for the bone hole without losing the correct adjustment against the cortical bone, thanks to the special conformation of the head. A metallic sheath, removable from the guide-tube, allows the insertion of the screw (whose head has a wider diameter than the body) without removing the handle of the probe from its leaning point against the cortical bone (corresponding to the hole of the nail).
- 34 It is possible to obtain a three-dimensional image by setting the Z axis too (orthogonal to X and Y axes), determining absolute certain coincidence of the hole with the magnetic field, displayed on the screen by 2 coincident light signals. Such axis (Z) is obtained joining the cylinder containing the 4 magnetic iron sensor at 90° with another cylinder with 4 sensor at 90° but with a 45° rotation in relation to the first ones. The double evaluation allow the measurement of the magnetic field variations even on a sagittal plane, defining therefore the precise point (the only one) of coincidence between the hole and the signal.
- 35 The research of the metallic object and the localization of the hole in the internal fixation procedure with an intramedullary nail accomplished in this way largely reduces the use of radiological appliances with all the risks connected with the x-ray exposure ("stochastic" effects: genetic, leukemogenetic and oncogenetic damages directly correlated to the absorbed dose). The system also reduces the operation time, allowing the identification and the centering of nail hole.
- 36 simultaneously.
- 37

RIVENDICATIONS

1. Use of the electromagnetic probe for the localization of metallic bodies in the organism.
2. Device able to invert the perception of the electromagnetic signal and to localize the hole (absence of metal) in metallic means used for osteosynthesis in Orthopaedics.
3. Use of a small magnet inserted into the hole (directly or by means of a rod) to enhance the precision of the hole localization by the electromagnetic probe.
4. Use of microelectromagnets able to send a high frequency modulatable electromagnetic signal, fixed to a rod sliding along the cavity of the fixation system, and which overlap the intramedullary nail hole; the signal is caught by the probe made of 4 magnetoresistive sensors, at crossways to increase precision.
5. Metallic (or teflon with preprinted circuit) rod containing 2 emitting coils, connected to a current generator at definite frequency and intensity, coincident with the holes, at definite distance, to be introduced into the medullary rod.
6. Magnetic device according to the preceding rivendications featured by the fact that it has been realised for the purposes and the uses above specified according to what has been described and illustrated.

1 / 7




FIG. 1




FIG. 2

2 / 7




FIG. 3






FIG. 4

3 / 7



4 / 7




FIG. 7

FIG. 8





FIG. 9








FIG.10

7 / 7

FIG. 11



INTERNATIONAL SEARCH REPORT

International Application No
PCT/IT 96/00179

A. CLASSIFICATION OF SUBJECT MATTER
IPC 6 A61B17/17

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)
IPC 6 A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	EP 0 097 401 A (CENTRE D'ETUDE DE L'ENERGIE NUCLEAIRE) 4 January 1984 see abstract see page 2, line 23 - line 35 ---	1,6
X	EP 0 589 592 A (ORTHOFIX) 30 March 1994 see abstract ---	1,6
X	EP 0 628 287 A (M.L.VALSECCHI) 14 December 1994 see abstract; figure 2 ---	3,6
X	WO 95 00085 A (S.B.HOLLSTIEN ET AL.) 5 January 1995 see abstract; figure 1 see page 12, line 14 - line 17 ---	3-6
	-/-	

☒ Further documents are listed in the continuation of box C.

☒ Patent family members are listed in annex.

* Special categories of cited documents :-

- *A* document defining the general state of the art which is not considered to be of particular relevance
- *E* earlier document but published on or after the international filing date
- *L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- *O* document referring to an oral disclosure, use, exhibition or other means
- *P* document published prior to the international filing date but later than the priority date claimed

T later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

X document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

Y document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

Z document member of the same patent family

Date of the actual completion of the international search

11 February 1997

Date of mailing of the international search report

19.02.97

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+ 31-70) 340-2040, Tx. 31 651 epo nl,
Fax (+ 31-70) 340-3016

Authorized officer

Nice, P

INTERNATIONAL SEARCH REPORT

International Application No
PCT/IT 96/00179

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT		
Category	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	EP 0 523 905 A (STRYKER) 20 January 1993 see abstract; figures 1,2 see column 5, line 52 - column 6, line 13 see column 7, line 2 - line 19 ---	3-6
P,X	DATABASE WPI Section PQ, Week 9625 Derwent Publications Ltd., London, GB; Class P31, AN 96-250140 XP002024975 & RU 2 046 377 C (Y.Y.REUTOV ET AL.) , 20 October 1995 see abstract -----	1

INTERNATIONAL SEARCH REPORT

International Application No

PCT/IT 96/00179

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
EP-A-0097401	04-01-84	LU-A- 84209 US-A- 4552134	07-03-84 12-11-85
EP-A-0589592	30-03-94	IT-B- 1257678 IT-B- 1262776 AU-B- 664113 AU-A- 4736293 CA-A- 2106529 CA-A, C 2106531 CN-A- 1084382 JP-A- 6277229 NO-A- 933360 US-A- 5433720 ZA-A- 9306955	01-02-96 04-07-96 02-11-95 31-03-94 23-03-94 23-03-94 30-03-94 04-10-94 23-03-94 18-07-95 15-06-95
EP-A-0628287	14-12-94	NONE	
WO-A-9500085	05-01-95	US-A- 5411503 AU-A- 7057194	02-05-95 17-01-95
EP-A-0523905	20-01-93	AT-T- 143580 CA-A- 2073266 DE-D- 69214219 JP-A- 6098894 US-A- 5584838	15-10-96 10-01-93 07-11-96 12-04-94 17-12-96